

CLEANING THE GM 200

From time to time or after exposure to sample material, the GM 200 needs to be cleaned comprehensively.

Remove knife and grinding jar.

Remove the black O-type sealing of the jar support. ①

Use a spatula or similar to unfix the black jar support and remove it. ②

Clean the GM 200 with a cloth dampened with water; make sure not to use too much water. ③

Dry the area and put the jar support in place again. ④

Fix the sealing. ⑤ ⑥

